



Actero™ Listeria Enrichment Media

Product Information

INTENDED USE:

Actero™ Listeria Enrichment Media (“*Actero™ Listeria*”) is a selective medium, specifically optimized for a 20-30 hours single-step recovery and enrichment of *Listeria* spp. from food and environmental samples.

PRINCIPLE OF OPERATION:

Actero™ Listeria Enrichment Media, “*Actero™ Listeria*”, combines nutritional components with additional ingredients that are necessary to selectively improve recovery and growth of *Listeria*. The selective agents present in the *Actero™ Listeria* have been optimized to efficiently target competing normal bacterial flora without affecting the growth of *Listeria* spp.

CONTENTS:

Actero™ Listeria Enrichment Media MediaBox™ 5 L (Cat# FCM-045) and *10 L* (Cat# FCM-046) contains respectively 5 L and 10 L of ready-to-use *Actero™ Listeria Enrichment Media*.

ADDITIONAL MATERIALS REQUIRED:

All Samples

1. Distilled/deionized water.
2. Sterile stomacher bags with and without filter.
3. Stomacher®.
4. Calibrated 10 µL plastic inoculating loops.
5. Rapid'L.mono agar plates (RLM).
6. Horse blood overlay agar (HL).
7. Modified Oxford agar plates (MOX).
8. Incubator: at $29 \pm 0.5^\circ\text{C}$, $32 \pm 0.5^\circ\text{C}$ and $35 \pm 2^\circ\text{C}$.
9. Regular laboratory equipment is also required.

Environmental Samples

1. Non-bactericidal sterile cellulose sampling sponges (8×4×0.3 cm) pre-moistened with neutralizing Dey-Engley buffer (D/E) (FoodChek Cat # FCLS-005).

PROCEDURE FOR MEDIABOX:

Materials not provided:

Ancillary tubing and connectors required can be purchased separately, see list of accessory tubing and connectors itemized later in the instructions for use.

Instructions:

Observe aseptic techniques. Stand the MediaBox with the Cap and dispensing tube at the top.



Set-up your dispensing tubing by connecting it to a pump or dilutor you use to control dispensing of the broth. Have your sterile connecting tubing with suitable connector ready to link to the MediaBox. Remove the sterile caps from both the dispensing tube attached to the MediaBox and the sterile connector stopper to your dispensing tubing. Connect the two pieces to permit flow of the broth into your dispensing tubing. Turn the MediaBox on its side with the dispensing cap and tubing toward the bottom of the MediaBox on the bench. Turn on your pump or dilutor and commence dispensing the broth into blender bags, bottles, tubes or other suitable vessels.

User Quality Control:

1. Examine initial dispensed broth from the MediaBox to confirm that the liquid is not cloudy, as this could indicate bacteria contamination in the MediaBox.
2. Inspect the MediaBox upon receipt for any signs of dampness on the outer box as this could indicate leakage of broth during transport.

PROCEDURE

Summary of Enrichment and Analysis of Samples Using Actero™ Listeria Enrichment Media

Sample Type	Sample Preparation	Analysis of Enriched Samples
Environmental Sample		
Food Contact Surface ➤ Stainless steel ➤ Plastic	<ol style="list-style-type: none"> 1. Swab a 100 cm² surface with a sponge (pre-moistened with D/E buffer) and keep it in a sterile bag at 4°C until tested. 2. Homogenize sponge with 90 mL pre-warmed (29°C or 32°C) Actero™ Listeria Enrichment Media. 3. Incubate at 29 ± 0.5°C for 24 h or at 32 ± 0.5°C for 28 h. 	<ul style="list-style-type: none"> ✓ By FoodChek™ Listeria spp. Assay ✓ By Actero™ Listeria Culture Method
Non Food Contact Surface ➤ Rubber ➤ Sealed concrete ➤ Ceramic	<ol style="list-style-type: none"> 1. Swab a 100 cm² surface with a sponge and keep it in a sterile bag at 4°C until tested. 2. Homogenize sponge with 90 mL pre-warmed (32°C) Actero™ Listeria Enrichment Media. 3. Incubate at 32 ± 0.5°C for 28 h. 	<ul style="list-style-type: none"> ✓ By FoodChek™ Listeria spp. Assay ✓ By Actero™ Listeria Culture Method
Non Food Contact Surface ➤ Sealed concrete	<ol style="list-style-type: none"> 1. Swab a 100 cm² surface with a sponge and keep it in a sterile bag at 4°C until tested. 2. Homogenize sponge with 90 mL pre-warmed (35°C) Actero™ Listeria Enrichment Media. 3. Incubate at 35 ± 2°C for 20-24 h. 	<ul style="list-style-type: none"> ✓ By DuPont™ BAX® System Real-Time PCR Assays for genus <i>Listeria</i> and <i>L. monocytogenes</i> ✓ By Actero™ Listeria Culture Method



Food Sample		
➤ Liquid whole eggs	<ol style="list-style-type: none"> Homogenize (30 s at 90 rpm) 25 g sample with 150 mL pre-warmed (32°C) Actero™ Listeria Enrichment Media. Incubate at 32 ± 0.5°C for 28 h. 	<ul style="list-style-type: none"> ✓ By FoodChek™ Listeria spp. Assay ✓ By Actero™ Listeria Culture Method
➤ Cold smoked salmon	<ol style="list-style-type: none"> Homogenize (30 s) 25 g sample with 150 mL pre-warmed (29°C) Actero™ Listeria Enrichment Media. Incubate at 29 ± 0.5°C for 28 h. 	<ul style="list-style-type: none"> ✓ By FoodChek™ Listeria spp. Assay ✓ By Actero™ Listeria Culture Method
	<ol style="list-style-type: none"> Homogenize (30 s) 25 g sample with 150 mL pre-warmed (35°C) Actero™ Listeria Enrichment Media. Incubate at 35 ± 2°C for 22-24 h. 	<ul style="list-style-type: none"> ✓ By DuPont™ BAX® System Real-Time PCR Assays for Genus <i>Listeria</i> and <i>L. monocytogenes</i> ✓ By Actero™ Listeria Culture Method
➤ Frozen cooked shrimp	<ol style="list-style-type: none"> Homogenize (30 s) 25 g sample with 150 mL pre-warmed (32°C) Actero™ Listeria Enrichment Media. Incubate at 32 ± 0.5°C for 28 h. 	<ul style="list-style-type: none"> ✓ By FoodChek™ Listeria spp. Assay ✓ By Actero™ Listeria Culture Method
	<ol style="list-style-type: none"> Homogenize (30 s) 25 g sample with 150 mL pre-warmed (35°C) Actero™ Listeria Enrichment Media. Incubate at 35 ± 2°C for 22-24 h. 	<ul style="list-style-type: none"> ✓ By DuPont™ BAX® System Real-Time PCR Assays for Genus <i>Listeria</i> and <i>L. monocytogenes</i> ✓ By Actero™ Listeria Culture Method
➤ Soft fresh cheese (Ricotta)	<ol style="list-style-type: none"> Homogenize (30 s) 25 g sample with 150 mL pre-warmed (32°C) Actero™ Listeria Enrichment Media. Incubate at 32 ± 0.5°C for 28 h. 	<ul style="list-style-type: none"> ✓ By FoodChek™ Listeria spp. Assay ✓ By Actero™ Listeria Culture Method
➤ Mexican-style cheese	<ol style="list-style-type: none"> Homogenize (30 s) 25 g sample with 150 mL pre-warmed (35°C) Actero™ Listeria Enrichment Media. Incubate at 35 ± 2°C for 22-24 h. 	<ul style="list-style-type: none"> ✓ By DuPont™ BAX® System Real-Time PCR Assays for Genus <i>Listeria</i> and <i>L. monocytogenes</i> ✓ By Actero™ Listeria Culture Method
➤ Pasteurized milk	<ol style="list-style-type: none"> Homogenize (30 s) 25 g sample with 150 mL pre-warmed (32°C) Actero™ Listeria Enrichment Media. Incubate at 32 ± 0.5°C for 28 h. 	<ul style="list-style-type: none"> ✓ By Actero™ Listeria Culture Method
➤ Bagged fresh spinach	<ol style="list-style-type: none"> Homogenize (30 s) 25 g sample with 150 mL pre-warmed (35°C) Actero™ Listeria Enrichment Media. Incubate at 35 ± 2°C for 22-24 h. 	<ul style="list-style-type: none"> ✓ By DuPont™ BAX® System Real-Time PCR Assays for Genus <i>Listeria</i> and <i>L. monocytogenes</i> ✓ By Actero™ Listeria Culture Method
➤ Hot dog sausage	<ol style="list-style-type: none"> Homogenize (30 s) 125 g sample with 750 mL pre-warmed (35°C) Actero™ Listeria Enrichment Media. Incubate at 35 ± 2°C for 26-28 h. 	<ul style="list-style-type: none"> ✓ By DuPont™ BAX® System Real-Time PCR Assays for Genus <i>Listeria</i> and <i>L. monocytogenes</i> ✓ By Actero™ Listeria Culture Method



Enrichment of Samples Using Actero™ Listeria Enrichment Media

NOTE: The sample preparation depends on the type and size of the sample. Thus, the protocol to prepare the sample should be chosen in function of those conditions. However, all the other steps are the same.

Environmental Surface Sample Preparation

1. Use a non-bactericidal sterile cellulose sampling sponge (8×4×0.3 cm) that has been pre-moistened with D/E buffer.
2. Swab the surface to be tested with one side of the swab(s) in a horizontal direction (approximately 10 cm), and with the other side in a vertical direction (approximately 10 cm) back and forth (one stroke back and one stroke forward) to swab the entire area of 100 cm².
3. Place each surface sampled sponge in a sterile sample bag, and keep at 4 ± 2°C until it is ready for testing. Analyze sample units as soon as possible after their receipt in the laboratory.
4. When ready to test, pre-warm the prepared *Actero™ Listeria Enrichment Media* at **29, 32 or 35°C** depending on the detection method to be used.
5. Add **90 mL** of the pre-warmed *Actero™ Listeria Enrichment Media* to each sponge sample in its sample bag.
6. Homogenize the sample for **30 seconds** in a Stomacher® 400 circulator. Hand mixing is an acceptable alternative for stomaching. To hand mix, massage each sponge that is in the sealed bag for approximately one minute.

Enrichment of Environmental Surface Samples Using Actero™ Listeria Enrichment Media

Actero™ Listeria Culture Method or FoodChek™ Listeria spp. assay

Food Contact Surface Samples (Stainless Steel, Plastic)

Close the bags and incubate the samples in an incubator for **24 h at 29 ± 0.5°C** or for **28 h at 32 ± 0.5°C**.

Non-Food Contact Surface Samples (Ceramic, Sealed Concrete, Rubber)

Close the bags and incubate the samples in an incubator for **28 h at 32 ± 0.5°C**.

DuPont™ BAX® System Real-Time PCR Assays for Genus *Listeria* and for *Listeria monocytogenes* or Actero™ Listeria Culture Method

Environmental Surface Samples (Stainless Steel, Plastic, Sealed Concrete)

Close the bags and incubate the samples in an incubator for **20-24 h at 35 ± 2°C**.

Note: Adherence to temperature is important for accurate results.

Food Sample Preparation

Liquid Whole Eggs (25 g)

Actero™ Listeria Culture Method or FoodChek™ Listeria spp. assay

1. Add **150 mL** of the pre-warmed (**32°C**) *Actero™ Listeria Enrichment Media* to each **25 g** sample in a **non-filtered** Stomacher® sample bag.
2. Homogenize the sample for **30 seconds at 90 rpm** in a Stomacher® 400 circulator or equivalent.
3. For the enrichment phase, close the bags and incubate the samples in an incubator for **28 h at 32 ± 0.5°C**. Adherence to temperature is important for accurate results.



Cold Smoked Salmon (25 g)

Actero™ Listeria Culture Method or FoodChek™ Listeria spp. assay

1. Add **150 mL** of the pre-warmed (**29°C**) *Actero™ Listeria Enrichment Media* to each **25 g** sample in a **filtered Stomacher®** sample bag.
2. Homogenize the sample for **30 seconds** in a Stomacher® 400 circulator or equivalent.
3. For the enrichment phase, close the bags and incubate the samples in an incubator for **28 h at 29 ± 0.5°C**. Adherence to temperature is important for accurate results.

DuPont™ BAX® System Real-Time PCR Assays for Genus *Listeria* and for *Listeria monocytogenes* or Actero™ Listeria Culture Method

1. Add **150 mL** of the pre-warmed (**35°C**) *Actero™ Listeria Enrichment Media* to each **25 g** sample in a **filtered Stomacher®** sample bag.
2. Homogenize the sample for **30 seconds** in a Stomacher® 400 circulator or equivalent.
3. For the enrichment phase, close the bags and incubate the samples in an incubator for **22-24 h at 35 ± 2°C**.

Frozen Cooked Shrimps (25 g)

Actero™ Listeria Culture Method or FoodChek™ Listeria spp. assay

1. Add **150 mL** of the pre-warmed (**32°C**) *Actero™ Listeria Enrichment Media* to each **25 g** sample in a **filtered Stomacher®** sample bag.
2. Homogenize the sample for **30 seconds** in a Stomacher® 400 circulator or equivalent.
3. For the enrichment phase, close the bags and incubate the samples in an incubator for **28 h at 32 ± 0.5°C**. Adherence to temperature is important for accurate results.

DuPont™ BAX® System Real-Time PCR Assays for Genus *Listeria* and for *Listeria monocytogenes* or Actero™ Listeria Culture Method

1. Add **150 mL** of the pre-warmed (**35°C**) *Actero™ Listeria Enrichment Media* to each **25 g** sample in a **filtered Stomacher®** sample bag.
2. Homogenize the sample for **30 seconds** in a Stomacher® 400 circulator or equivalent.
3. For the enrichment phase, close the bags and incubate the samples in an incubator for **22-24 h at 35 ± 2°C**.

Soft Fresh Cheese Ricotta (25 g)

Actero™ Listeria Culture Method or FoodChek™ Listeria spp. assay

1. Add **150 mL** of the pre-warmed (**32°C**) *Actero™ Listeria Enrichment Media* to each **25 g** sample in a **filtered Stomacher®** sample bag.
2. Homogenize the sample for **30 seconds** in a Stomacher® 400 circulator or equivalent.
3. For the enrichment phase, close the bags and incubate the samples in an incubator for **28 h at 32 ± 0.5°C**. Adherence to temperature is important for accurate results.

Mexican-style cheese (25 g), Bagged Fresh Spinach (25 g)

DuPont™ BAX® System Real-Time PCR Assays for Genus *Listeria* and for *Listeria monocytogenes* or Actero™ Listeria Culture Method

1. Add **150 mL** of the pre-warmed (**35°C**) *Actero™ Listeria Enrichment Media* to each **25 g** sample in a **filtered Stomacher®** sample bag.
2. Homogenize the sample for **30 seconds** in a Stomacher® 400 circulator or equivalent.
3. For the enrichment phase, close the bags and incubate the samples in an incubator for **22-24 h at 35 ± 2°C**.



Pasteurized Milk (25 g)

Actero™ Listeria Culture Method

1. Add **150 mL** of the pre-warmed (**32°C**) *Actero™ Listeria Enrichment Media* to each **25 g** sample in a **non-filtered** Stomacher® sample bag.
2. Homogenize the sample for **30 seconds** in a Stomacher® 400 circulator or equivalent.
3. For the enrichment phase, close the bags and incubate the samples in an incubator for **26-30 h at 32 ± 0.5°C**. Adherence to temperature is important for accurate results.

Hot Dog Sausage (125 g)

DuPont™ BAX® System Real-Time PCR Assays for Genus *Listeria* and for *Listeria monocytogenes* or Actero™ Listeria Culture Method

1. Add **750 mL** of the pre-warmed (**35°C**) *Actero™ Listeria Enrichment Media* to each **125 g** sample in a **filtered** Stomacher® sample bag.
2. Homogenize the sample for **30 seconds** in a Stomacher® 3500 or equivalent.
3. For the enrichment phase, close the bags and incubate the samples in an incubator for **26-28 h at 35 ± 2°C**. Adherence to temperature is important for accurate results.

Analysis of Enriched Samples

Actero™ Listeria Culture Method – All samples

1. At the end of the enrichment phase, streak a loopful of an enriched sample over the surface of one MOX and one RLM agar plates. Incubate the streaked plates for 24-48 h at 35 ± 2°C in an incubator.
2. After the 24 h incubation, examine both the MOX and RLM agar plates for colonies with morphology typical for *Listeria* spp. with following re-examination after 48 h of total incubation.
3. Use a loop to contact a minimum of 20 (if available) suspect colonies presented in MOX agar plate and collectively streak for isolation on one or more HL agar plates. Incubate the streaked HL plates at 35 ± 2°C for 22 ± 4 h.
4. Confirm the presumptive colonies growing on HL agar plates as *Listeria* spp. as recommended in the U.S. Food and Drug Administration *Bacteriological Analytical Manual*, Chapter 10, Detection and Enumeration of *Listeria monocytogenes* (2011) or in the FSIS Microbiology Laboratory Guidebook Chapter 8.09 (2013).

Interpretation and Test Result Report

1. If no suspect colonies with morphology typical of *Listeria* spp. have been found on MOX (typical *Listeria* spp. are small and are surrounded by a darker zone) and/or RLM agar plates (*L. monocytogenes* colonies will appear blue (pale blue, grey-blue to dark blue) without a yellow halo, *L. ivanovii* colonies will appear blue-green with a yellow halo and other *Listeria* strains will be white, with or without a yellow halo) following a 48 ± 2 h of total incubation, the sample is considered as being **negative for the presence of *Listeria* spp.**
2. If the **suspect colonies** with morphology typical for *Listeria* spp. have been found on MOX and/or RLM agar plates after 48 h of total incubation, the sample should be considered as **potentially positive for *Listeria* spp.** The suspect colonies **must be confirmed** according to the U.S. Food and Drug Administration *Bacteriological Analytical Manual*, Chapter 10, Detection and Enumeration of *Listeria monocytogenes* (2011) or in the FSIS Microbiology Laboratory Guidebook Chapter 8.09 (2013).



FoodChek™ Listeria spp. Assay – All samples

Refer to the Analysis section of the package insert for the FoodChek™ Listeria spp. assay.

DuPont™ BAX® System Real-Time PCR Assays for Genus *Listeria* and for *Listeria monocytogenes* – Environmental surface samples

Refer to the Test Protocol section of the package insert for each respective assay.

PRODUCT STORAGE AND SHELF LIFE:

On receipt, store the MediaBox at room-temperature 15-25°C. The expiry date is indicated on the package.

DISPOSAL:

Dispose all materials used and the enrichment medium by autoclaving or according to approved practices. Ensure that all biohazard waste is disposed of according to local, municipal, provincial, state and/or federal regulations.

PRECAUTIONS:

Biosafety level 2 procedures should be exercised.

(BMBL, <http://www.cdc.gov/biosafety/publications/bmbl5/bmbl.pdf>).

Extreme care should be taken in handling test samples and enrichment broths. All enrichment broths may contain various pathogens whether they contain *Listeria spp.* or not. Moreover, those who are at the highest risk of serious illness like pregnant women, and potentially immunocompromised individuals should have limited access to laboratory rooms or areas where *L. monocytogenes* isolation or identification procedures are in progress.

TERMS AND CONDITIONS:

FoodChek Systems Inc. makes no representations and warranties concerning its products other than those stated herein. All Product(s) delivered hereunder by FoodChek Systems Inc., its affiliates or any other person on its behalf shall, at the time of delivery, be manufactured to meet FoodChek Systems Inc.'s specifications and all applicable laws. All other terms, conditions and warranties, including any warranty of merchantability, quality, fitness or suitability for a particular or intended purpose, implied by common law or statute, (implied warranties) are expressly excluded.

CATALOGUE NUMBER:

FCM-045: Actero™ Listeria Enrichment MediaBox™ 5 L

FCM-046: Actero™ Listeria Enrichment MediaBox™ 10 L

FOR FURTHER INFORMATION PLEASE CONTACT:

FoodChek Systems Inc.
Suite 450, 1414 – 8 St. S.W.
Calgary, Alberta, Canada
T2R 1J6
Tel: 1-877-298-0208



Actero™

FoodChek™

Food safety, **simplified.**

